

ANEEXURE-I

Amendments in N.I.T for FUNCTION GENERATOR, Tendered Ref. No. M2/13/6000001793/VK1				
Clause No.	As per N.I.T		To be amended to	
	DESCRIPTION FOR BHEL REQUIREMENT	SPECIFIED BY	DESCRIPTION FOR BHEL REQUIREMENT	SPECIFIED BY
4.7	Amplitude Resolution	16 bits	Amplitude Resolution	14 bits or higher
4.8	Sampling Rate	16 bit 250 MSa/s	Sampling Rate	14 bit or higher at 250 MSa/s