

Technical Specification for Enquiry No. E-7130551

1. MULTIPARAMETER MONITOR:

DISCRIPTION:-

AC INPUT 100 TO 300V, OPERATING TEMPRATURE 0 DEGREE °C TO 50 DEGREE °C, 800X600 PIXELS, 8 CHENNEL MONITORS, GRADED AUDIO AND VISUAL ALARMS, ECG, RESPIRATION, DUAL TEMPRATURE, PULSE OXYMETRY, NON INVASIVE BLOOD PRESSURE, OXY CRG, DUAL IBP, CVP.

QTY.:- 01 no.

2. ECG MACHINE

DISCRIPTION:-

SIMULTANEOUS 3 CHANNEL ECG RECORDING WITH 12 LEAD SIMULTANEOUS ACQUISITIONS WITH AUTO SUMMARY. BUILT IN ECG PARAMETERS MEASURMENT AND INTERPRETATION, WIDE SCREEN LCD, HIGH PERFORMANCE LI-OIN RECHARGABLE BATTERY, POWER CAN BE USED BOTH ON POWER LINE AS WELL AS BATTERY.

QTY.:- 02 nos.

- A. **QUALIFYING CRITERIA:** - vendor should have supplied at least 3 – 4 machines earlier to other Hospital/Institution in last 2 - 3 years. Performance certificate of the previously supplied machine is also required.
- B. **Delivery Required in Dec-13**
- C. Erection & Commissioning is required within 30 days from the date of our call for commissioning. Please note that as per our requirement E&C charges should be equal or more than 2% of basic value of item. Please ensure that your quoted E&C charges is as per our requirement. If not, please confirm that payment as per clause no. 8.4 are acceptable to you.