

SPECIFICATION FOR CLAMP METER

Specification

Jaw opening	: 33 mm
AC Current range	: 1000 A
Accuracy	: 2 %
AC Voltage ranges	: 600 VAC
Accuracy	: 2.5 %
Resistance	: > 40 M Ω
Accuracy	: 2.5%
Frequency range	: 500 Hz
Continuity check	: audible at < 50 Ω
Display	: 3 ½ digits
Accessories	: Test leads to suit 600 VAC : carrying case
Battery	: cells with auto off function.

Kindly note:

1. The instrument offered can be better than the above specifications
2. Should have service facility in India with factory trained engineers to take care of the equipment supplied To reduce the down time.
3. The produce catalog should contain only the product offered in depth preferably highlighting the model Offered.
4. Two sets of product catalogue to be enclosed with the offer in original.
5. Offer without original catalog will be rejected.
6. Vendor should submit the offer with detailed specification of product offered. The offer given “ as per your specifications “ will be rejected.