


7

Pre-Qualifying Conditions for Tender No.-C/7029/11/6539 T

1. The Manufacturer should have standing of manufacturing the equipment for minimum 5 years.
2. The Manufacturer should furnish the list of customers where same model of equipment has been commissioned in last 5 years from the date of opening of the Tender.
3. The vendor should furnish satisfactory performance certificate from one customer where same model has been installed and has been working satisfactorily for last at least 1 year from the date of opening of the Tender. Also furnish the details of customer/company where the equipment has been installed with contact address and telephone no. of the customer.
4. The vendor should ensure regular after sales service including spares for at least five years after expiring of Warranty period.


27/3/12
Dr. S.K. Gupta
AGM (Surgery)

4 6

Technical specifications for Video Rectoscopy

1. Video Rectoscopy should have minimum 100W Halogen Lamp Light with spare Halogen Lamp
2. System should work on variable voltage
3. Light Intensity for illumination should be adjustable
4. Image Sensor should be HD-CCD system
5. Should have Zooming facility
6. Multiple display upto at least 8 frames
7. Should have flash memory for image storage
8. Signal processor should be digital & with PAL/NTSC system
9. Should have auto white balance
10. Should ~~be~~ have software for storing & printing images
11. Camera Cable of sufficient Length (2 metres)
12. Should come with standard accessories

- a) Probe - 1
- b) Metal Rectal Adaptor - 1
- c) Proctoscope - 1+1
- d) Sigmoidoscope - 1+1
- e) Air Pump Double Bellow - 1
- f) Foot switch - 1
- g) High ~~quality~~ T.V. Monitor at least 19" or more - 1
- h) Memory card - 1
- i) Cables & wires -
- j) Trolley for the system - 1
- k) Carrying case for Video Proctoscope - 1

14. one year warranty after successful commissioning.
15. Voltage stabilizer should be supplied.
16. Operating manual should be supplied.

Sum