

Technical Specification of enquiry no. E7130565

Description of item

AUTO REFRACTOMETER :- MAKE:- TOPCON/NIDEK/ HUMPHREY

Specification:- Sphere Range:- -25D to 22D (0.12D/ 0.25D)

Cylinder Range:- OD to +-10D(0.12D/ 0.25D) Corneal Curvature radius 5.00 to 10.00mm.(0.01mm)

PD measurement Range 20 to 85mm in 1 mm step. Display 8" to 9" LCD screen Rotary Prism technology

QUALIFYING CRITERIA:- Vendor should have supplied at least 3-4 machines earlier to other Hospital / Institute in last 2- 3 years. Performance certificate of the previously supplied machine is also required.

Qty:- 01 No.