

Technical Specification of Enquiry no. E7140573

DISCRIPTION: - MULTI-PARA MONITOR

SPO2, ECG, NIBP, TEMP, WITH ALARM & ARRYTHMIA DETECTION.

BATTERY BACKUP 120MIN WITH AC DC - 15 V DC.

DISPLAY - 7 TFT WITH HIGH RESOLUTION. SPO2 SHOW 0 TO 100% ECG 3 LEAD WITH PULSE RANGE 30-250 ECG DISPLAY IN 6 SEC. NIBP WITH AUTOMATIC & OSCILOMETRIC WITH SBP, DBP & MBP SHOW. WITH ACESSARIES ECG LEAD, WITH CABLE NIBP CUFF ADULT, PAEDIATIC, & INFANT. 7" (17.8CM.)COLOUR TFT LCD DISPLAY.

QUALIFIYING CRITERIA: - Mfg. Company should be certified by ISO 9001, CE & FDA.

Vendor should have supplied at least 3-4 machines earlier to other Hospital / Institute in last 2- 3 years. Performance certificate of the previously supplied machine is also required.

Errection & commissioning charges assumed approximate 2% of basic value, and E & C period shell be within 30 days from our call for commissioning.

QTY:- 01 No.