

Technical Specification for Enquiry No. E-7130540

1. Auto Refractometer :

SPHERE RANGE:- -25D TO 22D (0.12D/0.25D)

CYLINDER RANGE:- OD TO \pm 10D(0.12D/0.25D)

CORNEAL CURVATURE RADIUS 5.00 TO 10.00MM (0.01MM)

PD MEASUREMENT RANGE 20-85MM IN 1 MM STEP.

- A. **QUALIFYING CRITERIA:-** vendor should have supplied at least 3 – 4 machines earlier to other Hospital/Institute in last 2 - 3 years. Performance certificate of the previously supplied machine is also required.
- B. **Qty:** 01 No.
- C. **Delivery Required in Dec-13**
- D. Erection & Commissioning is required within 30 days from the date of our call for commissioning.
- E. You have indicated E&C charges separately in your offer. Please note that as per our requirement E&C charges should be equal or more than 2% of basic value of item. Please confirm your quoted E&C charges is as per our requirement. If not, please confirm that payment as per clause no. 8.4 is acceptable to you.