

Technical Specification for Enquiry No. E-7130541

1. CARDIAC DEFIBRILLATOR:-

BPL/SIEMENS/PHILLIPS POWER SUPPLY (AC INPUT) 100 TO 230V, OPERATING TEMPERATURE 00 TO 40 DEGREE C, CURRENT CONTROLLED BIPHASIC, MORE THAN 300 JOULES, 320X240 PIXEL.

PRINT AUTOMATION TIME, DATE, HEART RATE, HR LIMIT, EVENT MARKER ECG PARAMETER, DEFEBRI LLATOR MODE, SELECTED & DELIVERED ENERGY, PATIENT IMPEDANCE & HOSPITAL NAME.

- A. **QUALIFYING CRITERIA:-** vendor should have supplied at least 3 – 4 machines earlier to other Hospital/Institution in last 2 - 3 years. Performance certificate of the previously supplied machine is also required.
- B. **Qty:** 01 No.
- C. **Delivery Required in Dec-13**
- D. Erection & Commissioning is required within 30 days from the date of our call for commissioning.
- E. You have indicated E&C charges separately in your offer. Please note that as per our requirement E&C charges should be equal or more than 2% of basic value of item. Please confirm your quoted E&C charges is as per our requirement. If not, please confirm that payment as per clause no. 8.4 is acceptable to you.