



N/36  
SPECIFICATIONS - OF SPINE SURGERY  
INSTRUMENTS;

(TENDER. NO: 2011/4840)

- ① SELF RETAINING SPINAL RETRACTORS :-
  - a). SMALL ( 9.5 cms - 10.5 cms ) → 2 ~~Nos~~ <sup>Nos</sup>;
  - b). MEDIUM ( 12.0 cms - 13.0 cms ) → 2 ~~Nos~~ <sup>Nos</sup>;
  - c). LARGE ( 16.0 cms - 17.0 cms ) → 2 ~~Nos~~ <sup>Nos</sup>;
  
- ② GILLESPIE'S RETRACTORS :-  
( 16.0 cms - 17.0 cms ) → 2 NOS;
  
- ③ DURA/ROOT RETRACTORS - STRAIGHT :-  
( KOENING - TYPE ) → 2 NOS;
  
- ④ LAMINOTOMY FORCEPS ( JANSEN TYPE )  
SINGLE ACTION :
  - ↳ 2.0mm ( 16.0 cms - 18.0 cms ) → 2 NOS.
  - ↳ 3.0mm ( 16.0 cms - 18.0 cms ) → 2 NOS.
  
- ⑤ KERRISON PUNCH ( 40° ) X 10.0 cms.
  - upward hold jaw width -
    - 2 mm : → 1 NOS.
    - 3 mm : → 1 NOS.
    - 5 mm : → 1 NOS.
  
  - downward hold jaw width -
    - 2 mm : → 1 NOS.
    - 3 mm : → 1 NOS.
    - 5 mm : → 1 NOS.

*D. Singhal*  
(DR. D. M. SINGHAL)

2/19

PRE QUALIFYING CONDITIONS for Tenders: 4340;  
(C/7029/11/4340T)

- ①. Vendor shall submit - "Authorisation - letter"  
From the equipment manufacturer;
- ②. Vendor should submit <sup>satisfactory</sup> performance certificate  
from at least one customer.
- ③. Contact details of the customer should be submitted.

*P. M. Singhal*

DR. P. M. SINGHAL

Sr. Specialist (OR&CP)