

AUTOREFRACTOMETER

Technical Specifications

Requirements:

1. Autorefractometer with standard accessories and Motorised-table

Specifications:

Measurement Range:	Sphere	-25.00 to +22.00 D
	Cylinder	0 to +/- 10.00 D
	Axis	0 to 180

Measurable min Pupil diameter: 2 mm

Auto shooting

Chart: Scenery chart

PD measurement range: 30 to 85 mm

Display: 5.2 inch color LCD

Printer: Built in thermal type line printer

Accessories: Interface Cable, Barcode Scanners, Model eye

Table: Motorised

Warranty/ Guarranty 1 year may be asked from the supplier from the date of commissioning.



(Dr. S. Das)

Sr. Consultant (Ophth.)

For Auto Refractometer

Pre-Qualifying Requisites

(22-12-2012)

1. Vendor shall submit authorization from the equipment manufacturer.
2. Vendor shall provide the name, address and contact details of the customer(s) to whom offered item/similar item has been supplied in the past.
3. BHEL may ask the vendor to arrange satisfactory performance certificate of the equipment from his customer.

Seem

~~Signature~~

~~Signature~~

to Das

~~Signature~~

~~Signature~~

~~Signature~~