

**ANNEXURE-I (SPECIFICATIONS)****TENDER ENQUIRY-15E0101 DT:06.01.2016 DUE ON:29.01.2016**

1	Machine Type	Ac Type High voltage Tester
2	Output Voltage	0 to 100 KV AC
3	Capacity	leakage Current 300mA Max on H.V. Side
4	Operating voltage	415V, Single phase, 50 Hz
5	Duty Cycle	5Min ON/ 15min OFF
6	Indications	Indications required for mains On, HV on with Buzzer, HV off, Stroboscopic light for HV.
7	Indication Meter	All Indication meter input/output should be digital meter.
8	Transformer Cooling	Oil cooled ( first time oil to be filled by supplier)
9	variable Reactor	With Motorized control
10	Perferred Make	Tesla/ Other reputed brand
11	Mobility	Portable/ Suitable for transportation with lifting arrangement
12	Warranty	Two year from the date of commissioning
13	Test Certificate	Required
14	Training & Demostartion	At BHEL CFP Rudrapur
15	Opration & Maintinace Manal	Required
16	Mandatory spares	Required
17	Delivery	FOR Rudrapur

मनोहर सिंह धर्मशक्तु  
वरिष्ठ अभियंता (विद्युत)  
बी.एच.ई.एल. रुद्रपुर  
उत्तराखण्ड - 263153

M.S. Dharmshaktu  
Sr. Engineer (E)  
BHEL, Rudrapur  
Uttarakhand-263153