

Annexure-C to Tender BAP/CAP/2014-15/OT-11


BHEL:BAP:RANIPET -6.
MEDICAL DEPARTMENT

BAP:MED:ST:2014-15

DATE:03-12-2014.

Sub: Fully Automated Chemistry analyzer Specifications:

SL NO	DESCRIPTION
01	200 Tests / hour
02	Two Probe
03	Built in QC Programme
04	Built in Printer with external printer soft ware is a must
05	Software provided
06	Washing system and Distil water plant included
07	CE / US SBA APPROVED
08	Warranty to be 24 months from the date of installation. Amc charges for 3 years after expiry of warranty period to be given separately in offer.
09	Operator manuals, other relevant manuals needed
10	Service back up essential
11	Other features if any.
12	OPEN SYSTEM Reagents needed for the instrument ..
13	To be installed at BHEL Ranipet.


Dr. Mrs. Varalakshmi Baskar,
Consultant / General Medicine
BHEL HOSPITAL, RANIPET.