


Phones: 040-2318 2173/2536 Fax : 009-040-23186037				 BHARAT HEAVY ELECTRICAL LTD. RAMACHANDRAPURAM HYDRABAD-502032 ANDHRA PRADESH, INDIA		ENQUIRY			
				Enq. No : M1A1P08430		Enq. Date	No. of Items	DUE DATE OF QUOTATION	
						29.02.2016	7	22.03.2016	
Enq. Sr.No	PR.NO.	Item no	Material Code	Description	Unit	Quantity	Delivery Schedule		
							Date	Quantity	
10	1600008430	10	HY8470032801	PRESSURE GAUGE Standard design Nominal Size: 40mm / 1.5" Mounting: Back Direct Mounting Range: 0-40 Bar (Dual Scale in PSI) Connection: 1/4" BSP Accuracy Class: ±2.5 % F.S.D Case & Bezel (Crimped): SS304 Bourdon & Socket: SS316L Dial: Aluminium with White quoting Numerals: Black for Bar & Red for PSI indication Pointer: Aluminium black coated Window: Hardened Glass Filling Liquid: Glycerin Blow of Disc: to be provided on top Test and calibration certificate to be provided Warranty for one year. If found defective during this period it is required to rectify or replace at free of cost.	EA	200.00	13.04.2016	200.00	
20	1600008430	20	HY8470032810	PRESSURE GAUGE Standard design Nominal Size: 63mm / 2.5" Mounting: Bottom Direct Mounting Range: 0-50 Bar (Dual Scale in PSI) Connection: 1/4" BSP Accuracy Class: ±1.6 % F.S.D Case & Bezel (Crimped): SS304 Bourdon & Socket: SS316L Dial: Aluminium with White quoting Numerals: Black for Bar & Red for PSI indication Pointer: Aluminium black coated Window: Hardened Glass Filling Liquid: Glycerin Blow of Disc: to be provided on top Test and calibration certificate to be provided Warranty for one year. If found defective during this period it is required to rectify or replace at free of cost.	EA	50.00	13.04.2016	50.00	

30	1600008430	30	HY8451122612	<p>PRESSURE GAUGE Standard design Nominal Size: 63mm / 2.5" Mounting: Bottom Direct Mounting Range: 0-75 Bar (Dual Scale in PSI) Connection: ½" BSP Accuracy Class: ±1.6 % F.S.D Case & Bezel (Crimped): SS304 Bourdon & Socket: SS316L Dial: Aluminium with White quoting Numerals: Black for Bar & Red for PSI indication Pointer: Aluminium black coated Window: Hardened Glass Filling Liquid: Glycerin Blow of Disc: to be provided on top Test and calibration certificate to be provided Waranty for one year. If found defective during this period it is required to rectify or replace at free of cost.</p>	EA	75.00	13.04.2016	75.00
40	1600008430	40	HY8420692611	<p>PRESSURE GAUGE Standard design Nominal Size: 63mm / 2.5" Mounting: Bottom Direct Mounting Range: 0-100 Bar (Dual Scale in PSI) Connection: ½" BSP Accuracy Class: ±1.6 % F.S.D Case & Bezel (Crimped): SS304 Bourdon & Socket: SS316L Dial: Aluminium with White quoting Numerals: Black for Bar & Red for PSI indication Pointer: Aluminium black coated Window: Hardened Glass Filling Liquid: Glycerin Blow of Disc: to be provided on top Test and calibration certificate to be provided Waranty for one year. If found defective during this period it is required to rectify or replace at free of cost.</p>	EA	100.00	13.04.2016	100.00
50	1600008430	50	HY8360192774	<p>PRESSURE GAUGE Standard design Nominal Size: 63mm / 2.5" Mounting: Bottom Direct Mounting Range: 0-200 Bar (Dual Scale in PSI) Connection: ½" BSP Accuracy Class: ±1.6 % F.S.D Case & Bezel (Crimped): SS304 Bourdon & Socket: SS316L Dial: Aluminium with White quoting Numerals: Black for Bar & Red for PSI indication Pointer: Aluminium black coated Window: Hardened Glass Filling Liquid: Glycerin Blow of Disc: to be provided on top Test and calibration certificate to be provided Waranty for one year. If found defective during this period it is required to rectify or replace at free of cost.</p>	EA	50.00	13.04.2016	50.00

60	1600008430	60	HY8360192766	<p>PRESSURE GAUGE Standard design Nominal Size: 63mm / 2.5" Mounting: Bottom Direct Mounting Range: 0-300 Bar (Dual Scale in PSI) Connection: ½" BSP Accuracy Class: ±1.6 % F.S.D Case & Bezel (Crimped): SS304 Bourdon & Socket: SS316L Dial: Aluminium with White quoting Numerals: Black for Bar & Red for PSI indication Pointer: Aluminium black coated Window: Hardened Glass Filling Liquid: Glycerin Blow of Disc: to be provided on top Test and calibration certificate to be provided Waranty for one year. If found defective during this period it is required to rectify or replace at free of cost.</p>	EA	50.00	13.04.2016	50.00
70	1600008430	70	HY8320066565	<p>PRESSURE GAUGE Standard design Nominal Size: 63mm / 2.5" Mounting: Bottom Direct Mounting Range: 0-400 Bar (Dual Scale in PSI) Connection: ½" BSP Accuracy Class: ±1.6 % F.S.D Case & Bezel (Crimped): SS304 Bourdon & Socket: SS316L Dial: Aluminium with White quoting Numerals: Black for Bar & Red for PSI indication Pointer: Aluminium black coated Window: Hardened Glass Filling Liquid: Glycerin Blow of Disc: to be provided on top Test and calibration certificate to be provided Waranty for one year. If found defective during this period it is required to rectify or replace at free of cost.</p>	EA	50.00	13.04.2016	50.00
<p>Remarks:</p> <ol style="list-style-type: none"> General Condition of contract (GCC) HY: MM: GCC Rev.00 is applicable. Instruction to Bidder (ITB) HY: MM: ITB rev.04 attached along with this enquiry shall be duly filled in and submitted with sign & seal to consider your offer. DESTINATION: FOR BHEL basis for Indigenous Vendors inclusive of Freight and Insurance charges. As per ITB for important vendors Tender is on Two Part Bid System i.e. Part-I is Technical Bid and Part-II is Price Bid. Part-I bid will be opened first. Part-II will be opened at later date after technically and commercially qualified Vendor's offers. <u>Suppliers shall send offer/ quotations to the following address:</u> To <u>The DGM/Purchase (co-ordination), Vendor Complex,</u> <u>Bharat Heavy Electrical Limited,</u> <u>Ramachandrapuram, Hyderabad-502032.</u> In Case of Emergency, Bidder can also submit offer through email at his own risk. The offer to be submitted in two parts. Technical offer to be submitted to <u>technicalbid@bhelhyd.co.in</u>, and price bid to be submitted to <u>pricebid@bhelhyd.co.in</u>. Interchanging the information in the mails may lead to rejection of the offer. Supplier shall have no claim on e-mail offers sent on any other e-mail ID. BHEL is no way responsible for non-receipt of offers sent thru email due to server break down / Internet failure / transmission error etc. In case of e-mail offers, the mail subject should contain Enquiry No., Due date and Supplier name. Supplier address including contact details shall be mentioned in the content of the mail. Without these details offer may be liable for rejection. The signed (with seal) copy of specification, if any, along with GCC and ITB shall be submitted in part-I of your offer. Offers will be evaluated on Destination Cost to BHEL on item wise. Unless otherwise stated, in case of any ambiguity in commercial terms acceptance in ITB and your offer, ITB will be considered and binding upon you. 					TEST CERTIFICATE REQ: YES GUARANTEE REQ: YES SAMPLE REQ: No	For and on-behalf of Bharat Heavy Electricals Limited . Tushar Lonare Sr Purchase Officer/CMM		