

### **TECHNICAL SPECIFICATIONS: NASAL ENDOSCOPE**

1. 0deg,4mm wide angle 175mm working length autoclavable sinuscope
2. 30 deg ,4MM WIDE ANGLE 175 MM WORKING LENGTH AUTOCLAVABLE SINUSCOPE
3. 90 DEG ,4-8 MM WIDE ANGLE 175 MM WORKING LENGTH AUTOCLAVABLE ENDOSCOPE
- 4)cold LIGHT SOURCE DOUBLE OUT LET
5. FIBRO OPTIC CABLE
6. HD CAMERA DIGITAL MODEL WITH COUPLER WIRE CORD AND POWER SUPPLY
7. VIDEO IMAGING SYSTEM FOR ENDOSCOPIC APPLICATION  
SPECIAL PROGRAMME FOR ENDOSCOPE
- 8- VIDEO PROCESSOR HD QUALITY WITH 2X ZOOM MAGNIFICATION
- 9- LCD COLOUR MONITOR 26 INCH FULLY COMPLIANT WITH MEDICAL SAFETY , HIGH RESOLUTION WITH NATURAL COLOUR REPRODUCTION , HIGH BRIGHTNESS PERFORMANCE
- 10-OPTICAL ADAPTER
- 11-PUSH CART
- 12-COMPUTER
- 13-COLOUR PRINTER
- 14-FOOT SWITCH
- 15- Endodigisoftware special
- 16- Vendor has to support the specifications with manufacturer's brochure failing which offer will be rejected. Vendor has to demonstrate the equipment at BHEL General Hospital, Hyderabad within specified time limit, if asked for; failing which offer will be rejected