

TECHNICAL SPECIFICATIONS: NASAL ENDOSCOPE

1. 0 deg,4mm WIDE ANGLE 175MM WORKING LENGTH AUTOCLAVABLE SINUSCOPE
2. 30 deg /45 deg ,4MM WIDE ANGLE 175 MM WORKING LENGTH AUTOCLAVABLE SINUSCOPE
3. 90 DEG ,4MM /8 MM WIDE ANGLE 175 MM WORKING LENGTH AUTOCLAVABLE ENDOSCOPE
- 4) COLD LIGHT SOURCE DOUBLE OUT LET
5. FIBER OPTIC CABLE
6. CAMERA DIGITAL MODEL WITH MORE THAN 400 K pixel RESOLUTION WITH COUPLER WIRE CORD AND POWER SUPPLY
7. VIDEO IMAGING SYSTEM FOR ENDOSCOPIC APPLICATION WITH SUITABLE SOFTWARE AND FOLLOWING COMPONENTS:
 - A - LCD COLOUR MONITOR
 - B -OPTICAL ADAPTER
 - c -PUSH CART
 - D- -COMPUTER
 - E -COLOUR PRINTER
 - F -FOOT SWITCH
- 8- VENDOR HAS TO SUPPORT THE SPECIFICATIONS WITH MANUFACTURER'S BROCHURE FAILING WHICH OFFER MAY BE REJECTED. VENDOR HAS TO DEMONSTRATE THE EQUIPMENT AT, HYDERABAD WITHIN SPECIFIED TIME LIMIT, IF ASKED FOR; FAILING WHICH OFFER MAY BE REJECTED.