


Pre – Qualifying Requisites

01. Vendor shall submit authorization from the equipment manufacturer.
02. Vendor shall provide the name, address and contact details of the customer (s) to whom offered item / similar item has been supplied in the past.
03. Vendor should submit to arrange satisfactory performance certificate of the equipment from his customer.


(Dr. A.K. Shukla)
Sr. Specialist (Dental)


TECHNICAL SPECIFICATION

Dental chair.

Ergonomically designed Electrically operated computer controlled Dental Chair with a trolley mount cart system dental treatment centre, single foot control concept, electrically powered motor / up and down movements programmable, LED indicator, operating light cuspidor unit, system for pure water.

A regged dental chair as specified.

01. Operating light – 25000 LUX with neutral illumination.
02. Three way syringe with two aerator handpieces.
03. Highhand low vaccum (dry and wet) motorized suction with flow controlled valve.
04. Electrical Micrometer 40,000 rpm / with straight and contra angle hand pieces.
05. Medical grade OIL free mono bloc 1.5 HP compressor.
06. Ultrasonic hand piece scalar (Satelac) with 4 tips no. 1, Endo tips and light weight air polisher.
07. A training of doctor and staff at the time of commissioning of the equipment operation and maintenance free of cost.
08. The system will be declared commissioned only after successful commissioning of complete system.
09. Zero backache pneumatic stool.
10. 24 months gurantee after successful commissioning of complete system.



(Dr. A.K. Shukla)
Sr. Specialist (Dental)