













Index
For tender documents (Technical)
Tender no B/4011/2015/3647V2

Content	Page no. of pdf file
Specification ST45008 Rev. 01	Pg. 2 to Pg. 7
Sample Quality Plan Format	Pg. 8

दिनांक एवं हस्ताक्षर SIGN & DATE		उत्पाद मानक PRODUCT STANDARD	ST45008 पृष्ठ 6 का 1 Page 1 of 6
SUPERSEDES INVENTORY NO.	Based on standard : ASME PTC 6-2004		
इसकी पूरी सेवा के अधिकारि द्वारा	TECHNICAL SPECIFICATION FOR CALIBRATED FLOW NOZZLES		
COPYRIGHT AND CONFIDENTIAL The information on this document is the property of Bharat Heavy Electricals Limited. It must not be used directly or indirectly in any way detrimental to the interest of the company.	1.0 GENERAL This specification calls for supply of calibrated long radius, low beta ratio throat tap flow nozzle assembly(ies) for the accurate determination of primary flow in the Performance Guaranteed test on the TG set of the project indicated in sheet-5.		
स्वत्वाधिकार एवं गोपनीय इस दस्तावेज में सभी सूचनाएँ भारत भारी विद्युत् भारी उद्योग लिमिटेड की संपत्ति हैं। इस दस्तावेज को बिना अनुमति के किसी भी प्रकार से प्रसारित करने से बचना चाहिए।	2.0 SCOPE Supply of calibrated flow nozzle assembly(ies) comprising primary flow element, flow straightener(s), end flanges, gaskets (including one extra set for future use), requisite zinc plated bolts, nuts, upstream and downstream pipe sections and suitable root valves for tap sets. Each of the four pressure tap sets shall be 90° apart and calibrated as per ASME PTC 6-2004.		
दिनांक एवं हस्ताक्षर SIGN & DATE 27-10-10	3.0 TECHNICAL REQUIREMENTS Long Radius, low beta ratio throat tap flow nozzle assembly is to be designed as per para 4.8.4 to 4.8.11 of "ASME PTC 6-2004", "ASME PTC 19.5 -Flow Measurement" alongwith data in Sheet No. 4. The overall dimensions of the flow meter(s) should match with those indicated in sketch in sheet No.5. Major technical requirements are reproduced as follows:		
इसकी पूरी सेवा के INVENTORY NO. P-6404	a) The flow nozzle(s) shall be designed, manufactured and calibrated as per ASME PTC 6-2004. b) The supplier should calibrate the nozzle(s) preferably by Weigh tank method with Reynolds number, Water Temperature & other flow conditions as close to test conditions as possible. When it is not possible to calibrate at the test Reynolds number, the calibration Reynolds number should be obtained in accordance with para 4.8.16 of ASME PTC 6-2004. While calibrating, the transition region from 1.0 million to 4.0 million should be established .It is recommended that the value of the coefficient be established at the highest Reynolds No. possible, to minimize the effects of the transition region (in accordance with para 4.8.13 of ASME PTC 6-2004). It is recommended that nozzles be sized to produce throat Reynolds Nos. beyond this range if possible to avoid effects of the Transition region as described in para 4.8.17, pg.37 of ASME PTC 6-2004. c) The supplier should get the flow nozzle calibrated only at a recognized facility having international repute. Calibration should be conducted on all the pairs of taps & preferably consist of at least 20 acceptable points over a wide range of Reynolds Nos. For each set of selected taps, the calibration curve shall be within 0.25% of the reference curve and shall have the same slope. The reference curve is given in fig. 4.10 & table 4.2 on pg. 37 of ASME PTC 6-2004. The calibration of the flow nozzle shall comply with the conditions listed in clause 4-8.14 and 4-8.15 on pg. 36 & 37 of ASME PTC 6-2004.		
दिनांक एवं हस्ताक्षर DATE & SIGNATURE	नाम NAME	अनुवादक TRANSLATED BY	दिनांक एवं हस्ताक्षर SIGNATURE & DATE
TSX MEMBER-PSC	BHARI CHAUDHARY VIRENDRA KUMAR	 	- -
QAX संकेत विभाग AGREED DEPT.	S.K. CHAUHAN नाम NAME	 दिनांक एवं हस्ताक्षर DATE & SIGNATURE	PREETAM SINGH PRABHAT SINGH  
REV.NO. 09/05/11	स्वीकृति APPROVED : S. BHATTACHARYA DGM (STE)	निर्माण PREPARED : STE	जारी ISSUED : STE -TC दिनांक DATE : 07.10.2010

	उत्पाद मानक PRODUCT STANDARD		ST45008	
			पृष्ठ 6 का 2 Page 2 of 6	
SUPERVISOR INSPECTION	<p>d) The flow nozzle(s) shall be made from corrosion resistant material with known thermal expansion coefficient. Its surface should be hydraulically smooth and shall be free from all burrs, scratches, imperfections or ripples. For additional requirements on the design and manufacture of the nozzle(s) and its pressure taps, refer paras. 5.1, 5.2, 5.3, 5.4 of ASME PTC 19.5-2004 Flow Measurement & paras 4.8.6 & 4.8.7 of ASME PTC 6-2004.</p>			
	<p>e) The flow straightener shall be of Perforated or tubed plate design with non uniform hole distribution (<i>also called Gallagher Straightener</i>) as shown in Fig 4.5 & described in Para 4.8.4 & Table 4.1 on Pg. 30 & 32 of ASME PTC 6-2004.</p>			
COPY RIGHT AND CONSENT, IAI <small>I hereby certify that the design and drawings are the property of IAI and shall remain confidential.</small>	<p>f) The bidder should confirm that in case the calibration curves are not within the limits prescribed in ASME PTC 6, they will manufacture fresh flow nozzle(s) and calibrate it to achieve the calibration regimes of ASME PTC 6 and supply the new piece(s) without any price / delivery implications.</p>			
	<p>g) The total length of the flow assembly including flow straighteners; upstream & downstream machined pipe length; flanges of size Nb450 CL400 RF Weld neck type as per ANSI B16.5 etc. is to be exactly as indicated in the sketch in sheet No.6 of this document. For ease of transportation and installation, it is proposed to provide an interlocking flange joint near the center of assembly. This joint to be opened for shipping after calibration.</p>			
DESIGN ENGINEER IAI	<p>h) The external surfaces of the assembly shall be suitably painted to avoid rusting due to sea transportation (if applicable) and storage. The end faces shall be provided with suitable covers and conservation shall be done by filling Nitrogen and blanking the nozzle at the ends to avoid rusting/damage of the pipe internals. The nozzle shall remain filled with nitrogen until installation for Performance Test. Open ends of upstream and downstream taps should be sealed with covers to prevent damages to thread. The nozzles shall be packed in suitable seaworthy crates to avoid damage during transport and temporary storage before use at site.</p>			
	<p>i) Serial No. of the nozzle, tag FP01 should be clearly indicated on a name plate permanently affixed to the assembly. Tap set identification mark(s) shall also be punched as well as painted on the assembly so that they can be easily identified by the customer at site. These should also be clearly indicated in the assembly drawing.</p>			
INVESTIGATOR IAI	<p>j) The nozzle should be placed in box with name plate facing upward. The box should have arrow mark indicating upper direction and should atleast carry name of supplier, S.No. Of nozzle, name of the project, package size, gross weight and net weight.</p>			
REV. NO. 01			जांचकर्ता WORKED BY PREETAM SINGH	 09/05/11
			परीक्षणकर्ता CHECKED BY PRABHAT SINGH	 09/05/11

दिनांक व हस्ताक्षर SIGN & DATE		उत्पाद मानक PRODUCT STANDARD	ST45008 पृष्ठ 6 का 3 Page 3 of 6
SUPERSTRESS INVENTORY NO	4.0 DOCUMENTS TO BE SUBMITTED WITH OFFER		
समीक्षा संख्या व अधिकारिता संख्या	1) A copy of the design calculation for the nozzle(s) indicating clearly the calculation formula & the various correction factors like Buoyancy factor, Density correction, scale expansion etc. the bidder intends to apply in the calculations. 2) A typical flow calibration curve for a flow nozzle supplied earlier as per ASME PTC 6 3) Preliminary drawing of nozzle assembly and nozzle cross sectional drawing. 4) A quality plan on prescribed format. 5) Details of Manufacturing, testing & calibration facilities. 6) Experience list of supply of flow nozzle for similar rating supercritical thermal sets.		
COPYRIGHT AND CONFIDENTIAL The information in this document is the property of Bharat Heavy Electrical Limited. It must not be used directly or indirectly in any way detrimental to the interest of the company.	5.0 DOCUMENTS TO BE SUBMITTED FOR DESIGN APPROVAL (after Placement of order)		
	1) Drawings of the flow section with details of different components, material details and dimensions. 2) Final flow calculation as per ASME PTC 6. 3) An assembly drawing showing the location and assembly of all major components including the location of the marks showing tap set Nos. 4) Quality plan.		
स्वामित्वकर्ता एवं गोपनीय इस दस्तावेज में दी गई सूचना केवल ही प्रयोगकर्ता के उपयोग के लिए है। इसका प्रसारण अन्य किसी व्यक्ति को बिना लिखित अनुमति के नहीं किया जा सकता है।	5.1 INSPECTION & TEST CERTIFICATES		
	The following minimum tests shall be carried out as per applicable standards. 3 copies of test certificates and 1 reproducible shall be submitted for all the tests listed below before packing and dispatch of nozzles. 1) Calibrations on nozzle as per ASME PTC 6. Copies of calibration for each tap set indicating nozzle/pipe diameter and calibration conditions, including curves and calculation data shall be submitted for review and acceptance. 2) Material tests for all major components (i.e.) Nozzle, Steel pipes, Flanges, fittings, etc 3) Magnetic particle or liquid penetrant examination for each pressure taps as per ASME Sect.V. 4) 100 % Radiographic examination of butt welds as per ASME Sect.VIII. 5) Hydrostatic pressure test of complete meter run at 1.5 times of design pressure for minimum 30 minutes with proof of test. 6) Surface finish measurement for Nozzles satisfying the requirements of ASME PTC 6 and ASME PTC 19.5 as stated in para 3(d) above.		
दिनांक व हस्ताक्षर SIGN & DATE 11/10/10	REV. NO. 01 09/05/11		
समीक्षा संख्या INVENTORY NO. P-6401			जांचकर्ता WORKED BY PREETAM SINGH  07/10/10 परीक्षक/जांचकर्ता CHECKED BY PRABHAT SINGH  07/10/10

PRESCRIBED FORMAT FOR QUALITY PLAN

S. No	COMPONENT OPERATION	CHARACTERISTICS	CLASSIFICATION	TYPE OF CHECK	QUANTUM	REFERENCE DOCUMENT	ACCEPTANCE NORMS	FORMAT OF RECORDS	AGENCY P/W/V			REMARKS
1	2	3	4	5	6	7	8	9	10	11	12	13

COPYRIGHT AND CONFIDENTIAL
 The information contained herein is the property of Bharat Heavy Electricals Limited. It must not be used directly or indirectly in any way without the written consent of the company.



INSTRUCTIONS FOR FILLING QUALITY PLAN

The quality plan shall include all the quality control measures and check adopted by the vendor to ensure that the material/ components/ assembly/ services supplied by the vendor meet /will meet the requirements as per specifications and good practices. They shall include all stages of operation such as materials, process, assembly, packing & dispatch. The following guidelines for filling the Quality plan may be noted:








- Column 1- Serial Number
- Column 2- Component / operation. The component and / or operation being checked shall be given here.
- Column 3-Characteristics - The characteristic being checked shall be given here (e.g.) Chemical composition, mechanical properties, leak tightness, surface defects, etc.
- Column 4- Category CR stands for critical characteristic affecting the safety of equipment and personnel. MA stands for major characteristic affecting performance, reduction in life, large down time.MI stands for minor characteristic affecting appearance.
- Column 5-Type of check: like chemical analysis, tensile testing, hydraulic test, visual examination, radiography, etc.
- Column 6- Quantum of check such as 100%, 10%, 1 per Heat, etc.
- Column 7- Reference documents - Documents such as Technical specifications, drawings, standard specifications (BS, ANSI, ASME, DIN, IS, etc.), procedure etc. according to which the checks are done.
- Column 8- Acceptance norms-Standards etc. according to which the acceptability or otherwise of the characteristics being checked is decided.
- Column 9- Format of Records - Formats, log sheets, reports, etc. in which the observations are recorded. Standard log sheets, reports, format, etc. of the vendor shall be numbered and such reference numbers shall be included here.
- Column 10-12 -Agency - The agency which performs the test/inspection shall be written in the sub-column P. The agency which witnesses the tests shall be written in the sub-column W & the agency which verifies the tests certificates/inspection records shall be written in the sub-column V.

The information contained herein is the property of Bharat Heavy Electricals Limited. It must not be used directly or indirectly in any way without the written consent of the company.


REV. NO. 01
 09/05/11

REV. NO. 01 09/05/11		जांचकर्ता WORKED BY PREETAM SINGH	 07/11/10
		परीक्षक/प्रकृतो CHECKED BY PRABHAT SINGH	 07/11/10

4

दिनांक एवं हस्ताक्षर SIGN & DATE		उत्पाद मानक PRODUCT STANDARD	ST45008 पृष्ठ 6 का 5 Page 5 of 6																
सुपरीसेडर INVENTORY NO	DATA SHEET FOR FLOW MEASURING DEVICE 660 MW THERMAL POWER PLANT																		
श्रेणी एवं विवरण CLASSIFICATION	1) SERVICE : CONDENSATE FLOW TO DEAERATOR 2) TAG NO. : FP01 3) QUANTITY : 1 NO. 4) LINE FLUID : CONDENSATE (DM WATER) 5) OPERATING DATA																		
COPYRIGHT AND CONFIDENTIAL The information on this document is the property of Bharat Heavy Electricals Limited. It must not be used directly or indirectly in any way detrimental to the interest of the company.	<table border="1" style="width:100%; border-collapse: collapse;"> <thead> <tr> <th></th> <th>100 % LOAD (Design Point)</th> <th>80 % LOAD</th> <th>MAXIMUM LOAD</th> </tr> </thead> <tbody> <tr> <td>FLOW (T /Hr)</td> <td>1463.661</td> <td>1182.861</td> <td>1610.982</td> </tr> <tr> <td>PRESSURE(Kg/cm² abs)</td> <td>12.11</td> <td>10.03</td> <td>12.45</td> </tr> <tr> <td>TEMPERATURE (°C)</td> <td>153.8</td> <td>147.1</td> <td>154.8</td> </tr> </tbody> </table>				100 % LOAD (Design Point)	80 % LOAD	MAXIMUM LOAD	FLOW (T /Hr)	1463.661	1182.861	1610.982	PRESSURE(Kg/cm ² abs)	12.11	10.03	12.45	TEMPERATURE (°C)	153.8	147.1	154.8
	100 % LOAD (Design Point)	80 % LOAD	MAXIMUM LOAD																
FLOW (T /Hr)	1463.661	1182.861	1610.982																
PRESSURE(Kg/cm ² abs)	12.11	10.03	12.45																
TEMPERATURE (°C)	153.8	147.1	154.8																
23/5/11	6) PIPE SIZE : O.D. = 457.00 mm : I.D. = 431.60 mm : Thickness = 12.70 mm 7) PIPE MATERIAL : SA106 Gr C. 8) PIPE STANDARD : ANSI B 36.10 9) TYPE OF FLOW ELEMENT : LOW BETA RATIO THROAT TAP NOZZLE 10) MATERIAL OF FLOW ELEMENT : 321 SS or Equivalent 11) TYPE OF MOUNTING : Flanged 12) TYPE OF TAPPING : Throat Taps 13) NO OF TAPPING POINTS : FOUR Pairs 14) TYPE OF FLOW STRAIGHTENER : Perforated or Tubed plate (As per Fig. 4.5, ASME PTC-6-2004) 15) GRADIENT : Horizontal. 16) ALLOWABLE PRESSURE LOSS : Approx 1.0 Kg/cm ² at 100 % load 17) DESIGN & CALCULATIONS : As per ASME PTC-6-2004 & ASME PTC 19.5-2004 Flow Measurement 18) CALIBRATION CURVES & DATA : Required for each pair of tapings. Data indicating Cd _{theo} & Cd _{practical} to be furnished. 19) DIFFERENTIAL AT 100% LOAD : Preferably 1.5 Kg/cm ² 20) INSTRUMENT RANGE : 0-2.5 Kg/cm ² Dp transmitter																		
दिनांक एवं हस्ताक्षर SIGN & DATE	श्रेणी एवं विवरण INVENTORY NO P-6401	REV. NO. 01	<table border="1" style="width:100%; border-collapse: collapse;"> <tr> <td style="width:15%; text-align: center;"> कार्यकर्ता WORKED BY </td> <td style="width:35%; text-align: center;"> PREETAM SINGH </td> <td style="width:15%; text-align: center;">  09/05/11 </td> </tr> <tr> <td style="text-align: center;"> परीक्षक/कार्यकर्ता CHECKED BY </td> <td style="text-align: center;"> PRABHAT SINGH </td> <td style="text-align: center;">  09/05/11 </td> </tr> </table>	कार्यकर्ता WORKED BY	PREETAM SINGH	 09/05/11	परीक्षक/कार्यकर्ता CHECKED BY	PRABHAT SINGH	 09/05/11										
कार्यकर्ता WORKED BY	PREETAM SINGH	 09/05/11																	
परीक्षक/कार्यकर्ता CHECKED BY	PRABHAT SINGH	 09/05/11																	

REV. NO. 01



WORKED BY
CHECKED BY

PREETAM SINGH
PRABHAT SINGH

09/05/11

3-rod flange

PRODUCT STANDARD

ST45008

Page 6 of 6

DESIGN DATA						
PIPE I.D.	PIPE THICKNESS	PIPE MATERIAL	DESIGN PRESSURE	TEST PRESSURE	DESIGN TEMP.	DESIGN CODE
457.00	12.70	SA 106 GrC	48 Kg/cm ²	720 Kg/cm ²	160 Deg C	ANSI B36.1

All Dimensions in mm

