

Technical Specifications

Non contact type temperature Gun



Quantity Required: 05 nos


1. Temperature measurement range: upto 900 °C
2. Accuracy $\pm 1\%$ of reading.
3. Response time: less than 1 sec.
4. Adjustable emissivity.
5. Facility /Feature of Laser guide.
6. Optical resolution should better than or equal : 50:1.
7. Provision for data logging.
8. Instrument should be handy and light weight
9. Proper carrying case
10. Traceable calibration certificate to National/ International standards.
11. Guarantee/Warranty certificate: minimum one year.

Technical Pre-Qualification criteria:



Capability of the bidder for having supplied past above item or equivalent is a pre requisite. For that either

- At least one PO copy in the name of the bidder along with proof of delivery and calibration certificate in last five years to be submitted by the bidder alternatively
OR
- Bidder has to submit PO copies in the name of any bidder for having supplied 04 nos. of such item during previous 02 years


PSSR 12/11/2014 PSWR



PSNR PSER
(A.K. Gupta) (Vikash Kumar)

PSTS


(D.V. Mohan)

12/11/14
(Anil Gupta)