





दिनांक एवं हस्ताक्षर SIGN & DATE		उत्पाद क्रय विनिर्देश				PIR:RPC-14029HWW1 MC:ZZPC02913011		
		PRODUCT PURCHASE SPECIFICATION				पृष्ठ Page	2 1	का of
सुपरसेडस INVENTORY NO.	Flue Gas Analyser							
सभी सूची संख्या को अपेक्षित करना है	<p>Flue Gas Analyser suitable for Industrial applications viz. Thermal Power Stations, Diesel Generator Sets and Heat Treatment Furnaces etc. The following parameters would be monitored:</p> <p>Continuous Analysis of : Applicable to various fuel types: Coal, Natural Gas, Heavy Fuel Oil, Light Oil.</p>							
COPYRIGHT AND CONFIDENTIAL The information on this document is the property of Bharat Heavy Electrical, Limited. It must not be used directly or indirectly in any way detrimental to the interest of the company.	1.1 O ₂ : 0 - 21% 1.2 CO : 0-5000 ppm 1.3 NO : 0-4000 ppm 1.4 NO ₂ : 0-500 ppm 1.5 SO ₂ : 0-4000ppm 1.6 CO ₂ : 0-20% (NDIR) 1.7 Hydrocarbons 0-5000 ppm 1.8 Air temperature up to 50 ⁰ C 1.9 Flue gas temperature up to 999 ⁰ C 2.0 1 m long sampling probe with 3 m hose 2.1 Fresh Air Purging 2.2 Excess Air 2.3 Combustion Efficiency 2.4 Battery Charger 2.5. RS 232 Computer Interface 2.6. Printer Inbuilt with printer rolls 2.7. Data logging 2.8. Calibration Certificate 2.9. Carrying bag 2.10. Dust filter elements (pack of 10 elements) 2.11. Moisture free elements (pack of 5 elements) 2.12. Printer paper roll (pack of 5) 2.13. Printer cartridge							
	स्वाधिकार एवं गोपनीय इस प्रतीक में दी गई सूचना भारत भारती इंजीनियरिंग की संपत्ति है इसका प्रसार एवं उपयोग अन्य से किसी भी तरह करना, जो कि कंपनी के हित में सीमित है, न किया जाए।							
	हस्ताक्षर एवं SIGN & DATE					नाम NAME	दिनांक एवं हस्ताक्षर SIGNATURE & DATE	
	सभी सूची संख्या INVENTORY NO.					निर्माणकर्ता PREPARED BY	R S Yadav	<i>R S Yadav</i> 30/08/14
						जांचकर्ता CHECKED BY	R S Yadav	<i>R S Yadav</i> 30/08/14
		सहमत विभाग AGREED DEPTT.	नाम NAME	दिनांक एवं हस्ताक्षर DATE & SIGNATURE	पर्यवेक्षणकर्ता SUPERVISED BY		R S Yadav	<i>R S Yadav</i> 30/08/14
						स्वीकृति APPROVED : Ambrish Goel	<i>Ambrish Goel</i>	
	REV.NO.			निर्माण PREPARED : R S Yadav	जारी ISSUED : Ambrish Goel	दिनांक DATE : 30/08/14		
	Dt.							

दिनांक एवं हस्ताक्षर SIGN & DATE		उत्पाद क्रय विनिर्देश PRODUCT PURCHASE SPECIFICATION	PIR:RPC-14029HWW1 MC:ZZPC02913011			
			पृष्ठ	2	का	2
			Page	2	of	2
SUPERSEDES INVENTORY NO.	<p>3.0 General Terms & Conditions</p> <p>3.1 Installation & Commissioning: The supplier shall install & commission the equipment at site. Supplier should demonstrate the performance of the Equipment and all its accessories to ensure compliance with complete specifications and parameters quoted by them to the satisfaction of BHEL/end user.</p> <p>3.2 Document:</p> <p>(1) One hard copy of O&M manual in English language giving complete technical information for operation and maintenance of the equipment including circuit diagram shall be supplied.</p> <p>(2) Calibration Certificate from recognized agency / laboratory.</p> <p>3.3 Guarantee: The equipment shall be guaranteed for a period of 18 months from the date of dispatch or 12 months from the date of Installation & Commissioning, whichever is earlier.</p> <p>3.4 Packing & Dispatch: The equipment / accessories shall be packed in a strong suitable water proof packing boxes capable of bearing road transit hazards. Packing box shall be properly identified and marked with BHEL's(India) Purchase Order Number.</p> <p>4.0. Cross referred Document: Nil</p>					
सामग्री सूची संख्या को अपडेट किया गया है						
COPYRIGHT AND CONFIDENTIAL The information on this documents is the property of Bharat Heavy Electrical Limited. It must not be used directly or indirectly in any way detrimental to the interest of the company						
स्वामित्व अधिकार एवं गोपनीय इस प्रलेख में दी गई सूचना भारत हेवी इलेक्ट्रिकल्स की संपत्ति है इसका प्रयोग एवं प्रसारण के बिना भारत हेवी इलेक्ट्रिकल्स की लिमिटेड की अनुमति के बिना नहीं किया जा सकता है।						
हस्ताक्षर एवं दिनांक SIGN & DATE						
सामग्री सूची संख्या INVENTORY NO.	REV.NO.		निर्माणकर्ता PREPARED BY	R.S.Yadav		30/08/14
			जांचकर्ता CHECKED BY	Ambrish Goel		30/8/14

Indent No. 20142617 : Flue Gas Analyser

PRE-QUALIFYING REQUIREMENT (PQR) :

Only those vendors(OEMs) should quote who have supplied and commissioned at least one equipment of same size in the past ten years (on the date of opening of tender) and similar equipment is presently working satisfactorily for more than one year after commissioning at the supplied place (on the date of opening of tender).

The following information is to be submitted by the vendor about the companies where similar equipments have been supplied. This is required for all vendors for qualification of their offer.

- a. Name of the customer/company where referred system is installed
- b. Complete postal address of the customer.
- c. P.O.No., date, scope of supply and Month & year of commissioning.

Rajendra Singh Yadav
30/08/14

(Rajendra Singh Yadav)
Sr. Engineer (PCRI)

AGM (PCRI)

AGM
30/8/14