

7

Indent No.: C/6743/2014/1564T Rev 01
Specification No:-TB/BL-1/2014-15/03



SPECIFICATION FOR MULTI PHASE HIGH PERFORMANCE POWER ANALYZER

Purpose: - This Multi-phase high performance power analyzer is to be used for measurement of total power of 6pulse/12 pulse, 3-phase, dual winding VFD motor during testing of generators at Test Bed, Block-1.

Technical Specification:-

S.No.	Particulars	Requirement
1	Application	Measurement of total power of dual winding VFD motor. VFD motor can be operated by energizing single winding or both the windings simultaneously depending upon power requirement during generator testing. Each winding is 3-phase. <u>Motor Single Winding rating:</u> 7.1 MW, 4.1 kV, 1270 A, 3000 rpm, 6 pulse operation <u>Motor Dual Winding rating:</u> 12.0MW, 2x4.1 kV, 2x1068 A, 3000rpm, 12pulse operation The instrument should have sufficient channels for both voltage & current to give total power when motor is operated by energizing both the windings simultaneously.
2	Voltage Range	Upto 1000 Vpeak (Direct Input for High Accuracy)
3	Current Range	Upto 30 Amp. peak (Direct Input for High Accuracy)
4	Measurement Bandwidth Range	DC, 0.1 Hz. To 1MHz
5	Voltage / Current Frequency Range	5 MHz
6	Power Measurement Accuracy	+/- 0.2 % or better (overall accuracy)
7	Measurement Parameters	Voltage, Current, Active Power, Reactive Power, Apparent Power, Power Factor, Phase Angle, Frequency, Peak Voltage, Peak Current, Crest Factor, Integration, (WH, Ah, VARH, VAH)
8	Display	TFT/LCD Color
9	Harmonic Measurement Function	Instrument should have the facility to measure Harmonic Measurements in Voltage, Current, Individual Harmonic Levels, Harmonic Distortion Rate, Individual Harmonic Content for the specified Input Channels/Element up to minimum of 100th Order.
10	Display Update Rate	0.5 sec or better
11	Power Supply	220 VAC, 50 Hz.

[Signature]
10/7/14

[Signature]
10.7.14

[Signature]
10/7/14

Page 1 of 2

अनुपम चन्द
 उप महाप्रबन्धक (उत्पादन) ब्लॉक-1
 बी.एच.ई.एल., हीप, हरिद्वार

सुरेश कुमार भारद्वाज/Suresh Kumar Bhardwaj
 अपर महाप्रबन्धक (ब्लॉक-1)/AGM (Block-1)
 बी.एच.ई.एल., हीप हरिद्वार/BHEL, HEPP, Haridwar

Indent No.: C/6743/2014/1564T Rev 01
Specification No:-TB/BL-1/2014-15/03



SPECIFICATION FOR MULTI PHASE HIGH PERFORMANCE POWER ANALYZER

12	Demonstration of performance of power analyser at BHEL works	Vendor to demonstrate performance of power analyser at BHEL works during Generator testing. The value of power shall be compared with PMD (Power measuring device) currently being used at BHEL works for power measurement.
13	Additional Standard Features in High Performance Power Analyzer	<ul style="list-style-type: none"> • Standard USB Communication Interface for Data Logging to PC with requisite software. • Ethernet Communication for Remote Connection • Harmonic Data Values in the form of Bar Chart etc. should be available in the Software supplied with the instrument • Instrument should have facility for entering CT, PT & shunt ratio in order to facilitate direct display of actual value.
14	Warranty Certificate	Warranty Certificate required for minimum of 12 months from the date of dispatch against any manufacturing defects.
15	Calibration Certificate	Calibration Certificate document traceable to National accredited Lab is required
16	Standard accessories	<ul style="list-style-type: none"> • Voltage Lead Set (one per input module) • Suitable power cord • CD containing all relevant documentation and translated user manuals • USB Host - to - Device interface cable

S. S. Sane
10/7/14

S. B. B.
10/7/14

सुरेश कुमार भारद्वाज/Suresh Kumar Bhardwaj
 अपर महाप्रबन्धक (ब्लॉक-1)/AGM (Block-1)
 वी.एच.ई.एल., हीप हरिद्वार/BHEL, HEEP, Haridwar

Shand
10.7.14

अनुपम चन्द
 उप महाप्रबन्धक (उत्पादन) ब्लॉक-1
 वी.एच.ई.एल., हीप, हरिद्वार