


44

दिनांक एवं तिथि SIGN & DATE		उत्पाद क्रय विनिर्देश		PIR:RPC-14032HWW1 MC:ZZPC03214010	
		<b>PRODUCT PURCHASE SPECIFICATION</b>		पृष्ठ 2 Page 1	का 1 of 2

**6-DIGIT ELECTRONIC WEIGHING BALANCE**

**1.0 GENERAL:** To be used in the laboratory for weighing purposes, (6 Digit).

**2.0 TECHNICAL SPECIFICATIONS :**

2.1 Capacity : 5 gm Maximum

2.2 Readability : 1.0  $\mu$ g

2.3 Repeatability : 1.0  $\mu$ g

2.4 Display: LCD Display Alpha Numeric with backlite.

2.5 Calibration: Automatic system of Auto Internal Calibration, Complete range of Calibrated masses.

2.6 Self Auto Calibration

2.7 Response Time: 8-10 sec.

2.8 Units of Measurement: G, mg.




2.9 Standard Calibration Weights.

2.10 Operating Temp.: 18°C to 30°C

2.11 Inbuilt computer RS 232 interface

2.12 The equipment will work on 220 Volts AC  $\pm$  10%, 50 Hz  $\pm$  3%, single phase supply.

2.13 Detailed Printed Leaflet in Original of the model offered be supplied along with Quotation.

दिनांक एवं तिथि SIGN & DATE			नाम NAME		दिनांक एवं हस्ताक्षर SIGNATURE & DATE		
सुपरसेडस INVENTORY NO.			तैयारकर्ता PREPARED BY	R S Yadav	 14		
			जांचकर्ता CHECKED BY	R S Yadav	 14		
	सहमत विभाग AGREED DEPTT.	नाम NAME	दिनांक एवं हस्ताक्षर DATE & SIGNATURE	पर्यवेक्षणकर्ता SUPERVISED BY	R S Yadav	 14	
			स्वीकृति APPROVED : Ambrish Goel				
REV.NO.	00	तैयार PREPARED : R S Yadav		जारी ISSUED : Ambrish Goel	दिनांक DATE : 27/11/14		
Dt.							

**COPYRIGHT AND CONFIDENTIAL**  
 The information on this documents is the property of Bharat Heavy Electrical Limited. It must not be used directly or indirectly in any way detrimental to the interest of the company

**स्वाधिकार एवं गोपनीय**  
 इस दस्तावेज में दी गई सूचना भारत भारी बिजलीघर की संपत्ति है इसका प्रयोग एवं प्रसारण इस में किसी भी तरह प्रयोग, जो कि कंपनी के हित में अधिकारक हो न किया जाए।

43

SIGN & DATE उत्पाद क्रय विनिर्देश	PIR:RPC-14032HWW1 MC:ZZPC03214010			
	पृष्ठ Page	2 2	का of	2 2
SUPERSEDES INVENTORY NO. समीचीन संख्या की अधिसूचना संख्या है	<b>3.0 QUALITY ASSURANCE :</b> Calibration and Test certificate should be supplied along with the equipment.			
<b>COPYRIGHT AND CONFIDENTIAL</b> The information on this documents is the property of Bharat Heavy Electrical Limited. It must not be used directly or indirectly in any way detrimental to the interest of the company.	<b>4.0 General Terms &amp; Conditions</b>			
	<b>4.1 Installation &amp; Commissioning:</b> The supplier shall install & commission the equipment at PCRI, BHEL Haridwar. Supplier should demonstrate the performance of the Equipment and all its accessories to ensure compliance with complete specifications and parameters quoted by them to the satisfaction of BHEL/end user.			
	<b>4.2 Document:</b> ➤ One hard copy of O&M manual in English language giving complete technical information for operation and maintenance of the equipment including circuit diagram shall be supplied.			
	<b>4.3 Guarantee/Warranty:</b> The equipment shall be Guarantee/Warranty for a period of 18 months from the date of Dispatch or 12 months from the date of Installation & Commissioning, whichever is earlier.			
	<b>4.4 Packing &amp; Dispatch:</b> The equipment / accessories shall be packed in a strong suitable water proof packing boxes capable of bearing road transit hazards. Packing box shall be properly identified and marked with BHEL's(India) Purchase Order Number.			
स्वतंत्राधिकार एवं गोपनीय इस दस्तावेज में दी गई सूचना भारत में ही सर्वाधिकार की सुरक्षा के अंतर्गत है। इसका प्रयोग केवल भारत में ही किया जा सकता है। अन्य किसी भी देश में इसका प्रयोग, जो कि भारत में दिए गए अधिकारों के विरुद्ध है, गैरकानूनी है।	<b>5.0. Cross referred Document: Nil</b>			
SIGN & DATE				
INVENTORY NO.	REV.NO.	00	निर्माणांककता PREPARED BY	R.S.Yadav <i>Rajendra</i> 27/11/14
			जांचकर्ता CHECKED BY	Ambrish Goel <i>Ambrish</i>