

TECHNICAL SPECIFICATIONS: MULTI PARAMETER MONITOR

1-Electronic device To monitor ECG, SpO2.,NIBP, and Temperature with following features:

2-.Colour TFT display 8.4 “inches or bigger.

3 -Battery backup.

4 -72 hours or more measurement trends.

5- stat NIBP.

6 -SpO2.

7 -Suitable for Adult, pediatric and neonatal applications.

8 -Audio-visual alarm with preset limits.

9- Equipment should be supplied consisting of following:

a- Main unit

b -3 Lead ECG cable.

C- Adult and paediatric cuff and Hose.

D- Spo2 cable and probes

E -Temperature cable with probe.

10- Vendor has to support the specifications with manufacturer’s brochure failing which offer may be rejected. Vendor has to demonstrate the equipment at BHEL General Hospital, R.C.Puram, Hyderabad, within specified time limit, if asked for; failing which offer may be rejected.