

TECHNICAL SPECIFICATIONS: PHOTOTHERAPY UNIT

1- LED phototherapy with intensity upto 6-50microwatts/nM/cumm/mWatt, Adjustable intensity.

2-.Height adjustable and tiltable light unit.

3 -Digital time-totaller of LED usage time.

4 -Stainless steel tray.

5 –Provision for double surface phototherapy.

6- Vendor has to support the specifications with manufacturer's brochure failing which offer may be rejected. Vendor has to demonstrate the equipment at BHEL General Hospital, R.C.Puram, Hyderabad, within specified time limit, if asked for; failing which offer may be rejected.