

# SPECIFICATIONS

Item : Comprehensive Orthopaedic Hand Drill

## Technical Specifications :

01. Hand piece should be light weight i.e. less than 900 gms (without battery)
02. Hand piece should be "Battery Powered"
03. Hand piece should have variable speeds
04. Hand piece should have facility for "Finger Throttle" system
05. Hand piece should have facility for attachments for, tapping, drilling reaming, reciprocating and oscillating functions
06. Speed should be more than 1400 rpm and torque should be around 140 in / lbs
07. Should have option for Jacob's chuck and key less chuck
08. To provide separate cases for sterilization of autoclavable batteries
09. *warranty required for two years from date of commissioning;*

*J. Menzies*

### PRE QUALIFYING PERQUISITES

- Vender Should submit authorization from the equipment manufacturer. If vendor is not an authorized agent but original manufacture himself. He should submit a self-verified declaration on company's letter head stating that he is a manufacture with his signature and seal.
- Vendor shall provide the name, address and contact details of the customers to whom offered item / similar items has been supplied in the past. *- in last five (05) years; From the date of opening of tender;*
- Vender should furnish satisfactory performance certificate from one customer where same model has been installed and has been working satisfactory for last 6 months from the date of opening tender.

D. Singhal  
25/12/14  
(DR. I. M. SINGHAL)