

SN-19

**Boyle's ANAESTHESIA MACHINE .**

**TECHNICAL SPECIFICATION**

- 1 Anti- hypoxic device for gas ratio control
- 2 Anti- static coated 2 or 3 tubes Rota meter
- 3 N2O cut off on O2 failure
- 4 Stainless steel body with large anti-static castor wheel
- 5 Audio alarm for O2 failure
- 6 Half litter reservoir for storing O2 and emergency usage
- 7 Two auxiliary O2 outlet
- 8 Emergency oxygen flush at patient block
- 9 Patient safety blow-off valve when pressure exceeds 50cm/H2O
- 10 Selecta Tec assembly to mount TEC series vaporizer (Halothane & isoflurane)
- 11 Cylinder and pipeline connectivity
- 12 Bain and circle circuit connectivity
- 13 Circle absorber of ½ kg or 1 kg
- 14 Halothane vaporizer tec-7 ( Latest model) 1 no. with each machine
- 15 Isflurane vaporisor Tec 7 (latest model) 1 no with each machine.
- 16 Extended rear platform for two ten liter water capacity cylinders.
- 17 Pin index yokes ,two each for oxygen & N2O for 5 liter capacity.

18)


Warranty 4 yr from date *[Signature]*

of Commission

**Dr.A.K.Gupta**  
**Sr.Specialist (Anesthesia)**

## PRE QUALIFYING PERQUISITES

- Vender Should submit authorization from the equipment manufacturer. If vendor is not an authorized agent but original manufacture himself. He should submit a self-verified declaration on company's letter head stating that he is a manufacture with his signature and seal.
- Vendor shall provide the name, address and contact details of the customers to whom offered item / similar items has been supplied in the past.
- Vender should furnish satisfactory performance certificate from one customer where same model has been installed and has been working satisfactory for last 6 months from the date of opening tender.

  
Dr. Shamshad Ali  
MBBS, MD  
BHEL Ranipur  
Haridwar (U.K)