

HIGH PRESSURE STEAM AUTOCLAVE

TECHNICAL SPECIFICATION

Horizontal Autoclave in triple -walled(jacketed)

Technical data

Steam working pressure 1.26kgf/cm²

Operating voltage 440V, three phase, 50 Hz AC supply

Power for steam generation ; 9 kw

Capacity-Five drums-11x9" OR Three Drums-12x15"

Sterilization period: 45-50min

Radiant locking system

Cold water drainage system

Steam exhaust drainage system

Steam exhaust, double safety valve, steam releasing valve, gauge for pressure in side the chamber, heavy duty power plug socket provided with the unit .

Warranty 1 yr from date of commissioning.


Dr.A.K.Gupta
Sr.Specialist (Anesthesia)

PRE QUALIFYING PERQUISITES

- Vender Should submit authorization from the equipment manufacturer. If vendor is not an authorized agent but original manufacture himself. He should submit a self-verified declaration on company's letter head stating that he is a manufacture with his signature and seal.
- Vendor shall provide the name, address and contact details of the customers to whom offered item / similar items has been supplied in the past.
- Vender should furnish satisfactory performance certificate from one customer where same model has been installed and has been working satisfactory for last 6 months from the date of opening tender.



Dr. Ghanshad Ali
MBBS, MD
BHEL Ranipur
Haridwar (U.P.)