

SPECIFICATIONS

Item : Orthopaedic Rod Bender Cum Cutter

Specification


01. Instruments should be able to bend and/or cut rods of size from 3.0 mm to 6.5 mm
02. Instruments should be possessing "Built – in- Callibration" up to 6.5 mm
03. Instruments should have provision of long handles and provide better "cutting lever" and Bending Liver;
04. Instruments should be C.E. and ISO certified

as. One year warranty from date of commissioning;

D. M. Singh

PRE QUALIFYING PERQUISITES

- Vender Should submit authorization from the equipment manufacturer. If vendor is not an authorized agent but original manufacture himself. He should submit a self-verified declaration on company's letter head stating that he is a manufacture with his signature and seal.
- Vendor shall provide the name, address and contact details of the customers to whom offered item / similar items has been supplied in the past.
- Vender should furnish satisfactory performance certificate from one customer where same model has been installed and has been working satisfactory for last 6 months from the date of opening tender.


25/02/14
(DR. I. M. SINGHAL)